

SP

Notice of Allowability

Application No.

10/500,162

Examiner

John F. Ramirez

Applicant(s)

GENET ET AL.

Art Unit

3737

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 06/20/07.
2. ☒ The allowed claim(s) is/are 1-20.
3. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☒ All b) ☐ Some* c) ☐ None of the:
 1. ☒ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☐ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☐ Information Disclosure Statements (PTO/SB/08),
Paper No./Mail Date _____
4. ☐ Examiner's Comment Regarding Requirement for Deposit
of Biological Material
5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413),
Paper No./Mail Date _____
7. ☐ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____

DETAILED ACTION

Allowable Subject Matter

The following is an examiner's statement of reasons for allowance:

The prior art of record does not disclose or suggest, *inter alia*, an equipment for spectroscopic analysis of autofluorescence of a biological tissue comprising: an excitation source, a bundle constituted by a single optical fibre or a plurality of flexible optical fibres, means for injecting an excitation signal produced by said excitation source into said bundle according to a useful diameter corresponding to the excitation of the single fibre, all the optical fibres in the bundle or a specific sub-group, and a means for analyzing an emitted autofluorescence signal, an optical head at said bundle output, said optical head intended to be placed in contact with the biological tissue, said optical head being equipped with optical means adapted for converging the excitation signal coming out of said bundle into a subsurface analysis zone, the same optical fibre or fibres of said bundle having served for carrying the excitation signal being used for detecting the signal emitted by said subsurface analysis zone, and means placed upstream of the means for injecting provided for separating the excitation signal wavelength and the autofluorescence signal wavelength.

Concerning the section 103(a) rejection using Tsujita in view of Modell, the examiner agrees with applicant in that the cited references do not teach or suggest an equipment for spectroscopic analysis of autofluorescence of a biological tissue having an optical head being equipped with optical means adapted for converging the excitation signal coming out of said bundle into a subsurface analysis zone, the same

Art Unit: 3737

optical fibre or fibres of said bundle having served for carrying the excitation signal being used for detecting the signal emitted by said subsurface analysis zone, and means placed upstream of the means for injecting being moreover provided for separating the excitation signal wavelength and the autofluorescence signal wavelength.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."


Any inquiry concerning this communication or earlier communications from the examiner should be directed to John F. Ramirez whose telephone number is (571) 272-8685. The examiner can normally be reached on (Mon-Fri) 7:30 - 4:00 p.m.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Brian L. Casler can be reached on (571) 272-4956. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3737

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

JFR


BRIAN L. CASLER
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3737